

CONFORMITY CERTIFICATE

No. 201299322

Manufacturer: **IKO Science OÜ**
Laki str. 26
12915 Tallinn
ESTONIA

Product : **Biodetector**

Type: **BD500**

This conformity certificate confirms the conformity of the product with essential safety requirements of the following EC/EU New Approach Directives as amended:

2014/35/EU Low Voltage Directive
2014/30/EU EMC Directive

European harmonized standards used for the conformity assessment:

EN 61010-1:2010
EN 61326-1:2013
EN 61000-3-2:2014
EN 61000-3-3:2013

Other standards used for the conformity assessment:

EN 61010-1:2010/A1:2019

The certificate has been issued on the basis of the tests of the product type sample. The results are recorded in the Conformity assessment report No. 200500209/1 dated 04.11.2020



mark can be used only in the case of conformity assessment according to all relevant EC/EU Directives

This certificate is issued under the following conditions:

1. The certificate applies to the product type and its variations specified in the above mentioned Conformity Assessment report.
2. The production process/factory production control is not covered by this certificate.
3. The certificate does not imply that the certification body has performed any surveillance or control of the production process.
4. The manufacturer shall ensure the conformity of subsequent production items with the certified type.
5. Changes that may have an impact on maintaining conformity with the certification requirements may require confirmation of the validity of the certificate by demonstrating compliance with the conditions under which the certificate was issued or by conducting an additional evaluation.
6. The holder of this certificate must keep the conditions specified in the General Rules for Product Certification, which are freely available at www.tsu.eu

Issue date: 05.11.2020
Expiry date: 04.11.2023
Issue: 1




Ing. Dušan HANKO
Head of Product Certification Body

125785

CERTIFICATE

No. 0080/104/2020

Manufacturer: **IKO Science OÜ**
Laki str. 26
12915 Tallinn
ESTONIA

Product : **Biodetector**

Type: **BD500**

This certificate confirms the compliance of the product type characteristics with the technical requirements given in:

EN 60601-1-2:2015

The certificate has been issued on the basis of the tests of the product type sample. The results are recorded in the Conformity assessment report No. 200500209/2 dated 04.11.2020

This certificate is issued under the following conditions:

1. The certificate applies to the product type and its variations specified in the above mentioned Conformity Assessment report.
2. The production process/factory production control is not covered by this certificate.
3. The certificate does not imply that the certification body has performed any surveillance or control of the production process.
4. The manufacturer shall ensure the conformity of subsequent production items with the certified type.
5. Changes that may have an impact on maintaining conformity with the certification requirements may require confirmation of the validity of the certificate by demonstrating compliance with the conditions under which the certificate was issued or by conducting an additional evaluation.
6. The holder of this certificate must keep the conditions specified in the General Rules for Product Certification, which are freely available at www.tsu.eu

Issue date: 05.11.2020

Expiry date: 04.11.2023

Issue: 1




Ing. Dušan HANKO
Head of Product Certification Body

125782